

## Rabbit Type I Collagen Detection Kit

Catalog # 6016

*For Research Use Only - Not Human or Therapeutic Use*

### INTRODUCTION

Type I collagen is an interstitial fibrillar collagen consisting of two identical  $\alpha 1(I)$  chains and one  $\alpha 2(I)$  chain. It is the most abundant collagen type and is found in most connective tissues such as skin, bone, tendon, ligament, and heart. Many tissues contain heterotypic fibrils meaning that two or more distinct collagen types coexist. For example, most connective tissues (with the exception of bone) contain heterotypic fibrils of type I and III collagen, although type III collagen is a minor component.

The Rabbit Type I Collagen Detection Kit is designed to quantify the amount of type I collagen from rabbit cell and/or tissue culture or from tissue specimens by ELISA. All necessary reagents to measure 39 samples, in duplicate, are included in the kit. The linear range of this assay is approximately 0.16-10  $\mu\text{g/ml}$ .

Collagen Type	Antibody Specificity	
	Capture Antibody	Detection Antibody (Biotinylated)
Rabbit I	100%	70%
Denatured Rabbit I	2%	1%
Rabbit III	Not Determined	Not Determined
Mouse I	31%	0%
Chick I	0%	0%
Human I	100%	0%
Porcine I	100%	75%
Bovine I	100%	100%
Rat I	0%	100%

### ANTIBODY SPECIFICITY

This kit is not capable of measuring denatured rabbit type I collagen since a pair of IgG monoclonal antibodies to rabbit type I collagen used as the capture and detection antibodies are highly specific to the native conformation of type I collagen. In addition, samples of rabbit type I collagen mixed with other species of type I collagen cannot be used for this kit because the capture antibody cross-reacts to other species, such as human, porcine, bovine, and mouse type I collagen (see table above).

### KIT COMPONENTS

Item	Quantity	Amount	Storage
Rabbit Type I Collagen Standard	1 vial	150 $\mu\text{l}$ , 100 $\mu\text{g/ml}$	-20°C
Capture Antibody	1 vial	100 $\mu\text{l}$ , 1 mg/ml	-20°C
Detection Antibody	1 vial	Lyophilized	-20°C
Solution A - Capture Antibody Dilution Buffer	1 bottle	10 ml	-20°C
Solution B - Sample/Standard Dilution Buffer	1 bottle	50 ml	-20°C
Solution C - Detection Antibody Dilution Buffer	1 bottle	10 ml	-20°C
Solution D - Streptavidin Peroxidase Dilution Buffer	1 bottle	20 ml	-20°C
Streptavidin Peroxidase	2 vials	50 $\mu\text{l}$	-20°C
OPD	2 vials	Lyophilized	-20°C
Chromagen Dilution Buffer	1 bottle	20 ml	-20°C
Stop Solution - 2N Sulfuric Acid	1 bottle	10 ml	-20°C
Wash Buffer, 20X	1 bottle	50 ml	-20°C
ELISA Plate	1 each	96-well (8-well strips x 12)	-20°C

## NOTES ON PROCESSING SAMPLES

### Solubilization of Collagen

To determine the collagen content in cultured cell layers and tissues by ELISA, solubilization of collagen is required. To solubilize collagen from sample specimens, a limited digestion of tissue specimens with pepsin is recommended, although other neutral proteinases, such as pronase and papain are capable of solubilizing collagen. These proteinases only digest telopeptides located on both the N- and C-terminals of the collagen molecule, but are not capable of digesting the helical conformation region of the collagen molecule and intra- and inter-molecular cross-linkages.

The solubilization of collagen from tissues by a limited proteinase digestion (generally collagen/proteinase ratio is 100:1) depends on the types of tissues and the contents of intra- and inter-molecular cross-linkages. For example, bone and Achilles tendon are resistant to pepsin digestion, and only 10-20% of the collagen tissue will be solubilized. On the other hand, young calf skin collagen will be completely solubilized by pepsin digestion within 24-48 hours, but it takes 7-9 days to solubilize adult calfskin.

Proteinase resistant insoluble collagen might be solubilized by alkaline treatment. Suspend insoluble collagen in cold 0.1N NaOH solution containing 10% Na<sub>2</sub>SO<sub>4</sub> and 0.1M Amine such as Tris, and incubate at 4°C for 1-2 weeks. After treatment of collagen with alkaline, neutralize the pH to 5.0 with HCl, and then dilute it with 0.05M acetic acid or neutral buffer such as 0.1M Tris-0.3M NaCl, pH 7.5.

Therefore, the optimum solubilization condition for individual samples should be determined before processing samples. Collagen can be analyzed by 6% SDS-gel under non-reducing condition, using authentic type I collagen as a standard. If samples contain bands larger than the  $\gamma$ -chain (MW = 300 Kd), the samples must be further digested by pepsin or elastase. On the other hand, if smaller bands or smear bands are observed under the  $\alpha$ -chain (MW = 100 Kd), the samples might be over-digested. Therefore, it is critical to understand the biological and physico-chemical properties of individual collagen samples.

Tips on solubilization of collagen can be obtained from Chondrex customer service.

## NOTES BEFORE USING ASSAY

Note 1: It is recommended that the standard and samples be run in duplicate.

Note 2: Partially used reagents may be kept at -20°C.

Note 3: Crystals may form in the 20X wash buffer when stored at cold temperatures. If crystals have formed, it is necessary to warm the wash buffer by placing the bottle in warm water until crystals have dissolved completely.

Note 4: Measure exact volume of buffers using a serological pipette prior to diluting. Extra buffer is provided.

## ASSAY PROCEDURE

1. **Add Capture Antibody:** Dilute one vial of Capture Antibody with 10 ml of Capture Antibody Dilution Buffer (Solution A). Add 100  $\mu$ l of capture antibody solution to each well and incubate at 4°C overnight.
2. **Dilute Wash Buffer:** Dilute 50 ml of 20X wash buffer in 950 ml of distilled water (1X wash buffer). Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blot on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
3. **Prepare Standard Dilutions:** The recommended standard range is 0.16-10  $\mu$ g/ml. Prepare serial dilutions of the standard by mixing 50  $\mu$ l of 100  $\mu$ g/ml standard with 450  $\mu$ l of Sample/Standard Dilution Buffer (Solution B) - 10  $\mu$ g/ml. Then mix 250  $\mu$ l of the 10  $\mu$ g/ml standard with 250  $\mu$ l of Solution B - 5  $\mu$ g/ml. Then repeat this procedure to make five more serial dilutions of standard - 2.5, 1.25, 0.63, 0.32, and 0.16  $\mu$ g/ml solutions. The 100  $\mu$ g/ml standard stock may be stored at -20°C for use in a second assay. We recommend making fresh serial dilutions for each assay.
4. **Prepare Sample Dilutions:** Dilute tissue samples 1:1-1:1000 with Solution B depending on the estimated collagen content in the samples. Cell samples can be used without further dilution. However, if it is necessary, dilute cell samples 1:1-1:2 with Solution B.
5. **Add Standards and Samples:** Mix samples and standard tubes well. Add 100  $\mu$ l of Solution B (blank), standards and samples to appropriate wells. Incubate at room temperature for 2 hours.
6. **Wash:** Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blot on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
7. **Add Detection Antibody:** Dissolve one vial of Detection Antibody in 10 ml of Detection Antibody Dilution Buffer (Solution C). Add 100  $\mu$ l of detection antibody solution to each well and incubate at room temperature for 2 hours.
8. **Wash:** Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blot on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
9. **Add Streptavidin Peroxidase:** Dilute one vial of Streptavidin Peroxidase in 10 ml of Streptavidin Peroxidase Dilution Buffer (Solution D). Add 100  $\mu$ l of streptavidin peroxidase solution to each well and incubate at room temperature for 1 hour.
10. **Wash:** Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blot on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
11. **OPD:** Dissolve one vial of OPD in 10 ml of Chromagen Dilution Buffer just prior to use. Add 100  $\mu$ l of OPD solution to each well immediately after washing the plate. Incubate for 30 minutes at room temperature.
12. **Stop:** Add 50  $\mu$ l of 2N sulfuric acid (Stop Solution) to each well.
13. **Read Plate:** Read the OD values at 490 nm. If the OD values of samples are greater than the OD values of the highest standard, re-assay the samples at a higher dilution. A 630 nm filter can be used as a reference.

## CALCULATION OF RESULTS

1. Average the duplicate OD values for the blank, standards, and test samples.
2. Subtract the "blank" (B) values from the averaged OD values in step 1.
3. Plot the OD values of standards against the concentration of collagen ( $\mu\text{g/ml}$ ). Using a log/log plot will linearize the data. Figure 3 shows a representative experiment where the standard range is from 0.16 to 10  $\mu\text{g/ml}$ .
4. The  $\mu\text{g/ml}$  of rabbit type I collagen in test samples can be calculated using regression analysis.

Figure 3 - A typical standard curve

