

Mouse Anti-Peanut Ara h2 Antibody Subtype/Subclass ELISA Kits

Catalog # 3077, 3078, 3079, and 3080

For Research Use Only - Not Human or Therapeutic Use

PRODUCT SPECIFICATIONS

DESCRIPTION:	ELISA kit to quantify mouse anti-peanut Ara h2 antibodies
FORMAT:	96-well ELISA Plate with removeable strips
ASSAY TYPE:	Indirect ELISA
ASSAY TIME:	4.5 hours
STANDARD RANGE:	3077 (IgG) : 10 - 0.16 ng/ml 3078 (IgG2b) : 10 - 0.16 ng/ml 3079 (IgM) : 200 - 0.8 ng/ml 3080 (IgA) : 200 - 0.8 ng/ml
NUMBER OF SAMPLES:	Up to 40 (duplicate) samples/plate
SAMPLE TYPES:	Serum & Plasma
RECOMMENDED SAMPLE DILUTIONS:	1:100 (at least)
CHROMOGEN:	TMB (read at 450 nm)
STORAGE:	-20°C for 12 months
VALIDATION DATA:	3077: Intra-Assay (6.1-8.8%)/Inter-Assay (2.6-5.5%)/Spiking Test (99-108%) 3078: Intra-Assay (3.6-6.1%)/Inter-Assay (3.2-9.4%)/Spiking Test (106-109%) 3079: Intra-Assay (3.0-5.6%)/Inter-Assay (6.7-8.4%)/Spiking Test (100-110%) 3080: Intra-Assay (1.9-5.2%)/Inter-Assay (1.5-6.9%)/Spiking Test (94-103%)
NOTES:	N/A

Mouse Anti-Peanut Ara h2 Antibody Subtype/Subclass ELISA Kits

Catalog # 3077, 3078, 3079, and 3080

For Research Use Only - Not Human or Therapeutic Use

INTRODUCTION

Immediate hypersensitivity reactions to peanuts, an IgE-mediated food allergy, have been a major public health concern for many years, particularly in westernized countries where peanut allergies can persist into adulthood. For allergic patients, avoidance currently remains the only viable option (1).

Eleven potentially important peanut allergens have been identified. Ara h1, Ara h2, Ara h3, and Ara h6 have been designated the major peanut allergens. Ara h2 and Ara h6, two highly related 2S albumins, especially contribute to the development of allergic reactions (2).

Mouse peanut allergy models have been used to study the pathogenesis of the peanut allergy and to help develop new treatments. The mouse models can be induced by administration of crude peanut extract (CPE) or each purified Ara allergen and evaluated for the humoral immune responses such as serum anti-IgE and IgG antibodies against the allergen, T-cell mediated immune response associated cytokines levels, as well as body temperature and clinical signs of anaphylaxis. These factor changes observed in the disease models are useful for studying the efficacy of protective effects against the development of allergic reactions (3–9).

To evaluate the humoral immune response against Ara h2 in mouse allergy models, Chondrex, Inc. provides ELISA kits for assaying mouse anti-Ara h2 subtype and subclass antibodies including IgA, IgM, IgG, and IgG2b antibodies. Chondrex, Inc. also offers ELISA kits for assaying anti-CPE, ovalbumin, house dust mite, and gliadin antibody subtypes/subclasses as well as total immunoglobulin subtypes/subclasses. For more information, please visit www.chondrex.com or contact support@chondrex.com.

LIST OF MOUSE ANTI-PEANUT ANTIBODY SUBTYPE/SUBCLASS ELISA KITS

Kit	CPE Catalog #	Ara h2 Catalog #	Ara h3 Catalog #
Mouse Anti-Peanut IgG Antibody ELISA Kit	3056	3077	3082
Mouse Anti-Peanut IgG1 Antibody ELISA Kit	3057	Coming Soon!	3083
Mouse Anti-Peanut IgG2b Antibody ELISA Kit	3059	3078	Coming soon!
Mouse Anti-Peanut IgM Antibody ELISA Kit	3062	3079	Coming soon!
Mouse Anti-Peanut IgE Antibody ELISA Kit	3063	Coming soon!	3071
Mouse Anti-Peanut IgG2a Antibody ELISA Kit	3058	Coming soon!	Coming soon!
Mouse Anti-Peanut IgG3 Antibody ELISA Kit	3060	Coming soon!	Coming soon!
Mouse Anti-Peanut IgA Antibody ELISA Kit	3061	3080	Coming soon!

KIT COMPONENTS

	Item	Quantity	Amount	Storage
Standard	IgG (30771) – 10 ng IgG2b (30781) – 10 ng IgM (30791) – 200 ng IgA (30801) – 200 ng	1 vial	Lyophilized	-20°C
Secondary Antibody (peroxidase-conjugated polyclonal antibodies)	IgG (30233) IgG2b (30353) IgM (30173) IgA (30183)	2 vials	50 µl	-20°C
	Solution B - Blocking Buffer (30105)	1 bottle	10 ml	-20°C
	Solution C - Sample/Standard/Secondary Antibody Dilution Buffer (30106)	1 bottle	50 ml	-20°C
	TMB Solution (90023)	2 vials	0.2 ml	-20°C
	Chromogen Dilution Buffer (90022)	1 bottle	20 ml	-20°C
	Stop Solution - 2N Sulfuric Acid (9016)	1 bottle	10 ml	-20°C
	Wash Buffer, 20X (9005)	1 bottle	50 ml	-20°C
	Recombinant Ara h2 Coated ELISA Plate (Blue)	1 each	96-well (8-well strips x 12)	-20°C

ASSAY OUTLINE




PLATE MAPPING

Example of the Mouse Anti-Ara h2 IgG and IgG2b Antibody ELISA Kits

	1	2	3	4	5	6	7	8	9	10	11	12
A	10	10	S1	S1	S9	S9	S17	S17	S25	S25	S33	S33
B	5	5	S2	S2	S10	S10	S18	S18	S26	S26	S34	S34
C	2.5	2.5	S3	S3	S11	S11	S19	S19	S27	S27	S35	S35
D	1.25	1.25	S4	S4	S12	S12	S20	S20	S28	S28	S36	S36
E	0.63	0.63	S5	S5	S13	S13	S21	S21	S29	S29	S37	S37
F	0.31	0.31	S6	S6	S14	S14	S22	S22	S30	S30	S38	S38
G	0.16	0.16	S7	S7	S15	S15	S23	S23	S31	S31	S39	S39
H	B	B	S8	S8	S16	S16	S24	S24	S32	S32	S40	S40

Standards
Samples

NOTES BEFORE USING ASSAY

NOTE 1: It is recommended that the standard and samples be run in duplicate.

NOTE 2: Warm up all buffers to room temperature before use.

NOTE 3: Crystals may form in Wash Buffer, 20X when stored at cold temperatures. If crystals have formed, warm the wash buffer by placing the bottle in warm water until crystals are completely dissolved.

NOTE 4: Measure exact volume of buffers using a serological pipet, as extra buffer is provided.

NOTE 5: Cover the plate with plastic wrap or a plate sealer after each step to prevent evaporation from the outside wells of the plate.


NOTE 6: For partial reagent use, please see the assay protocol's corresponding step for the appropriate dilution ratio. For example, if the protocol dilutes 50 µl of a stock solution in 10 ml of buffer for 12 strips, then for 6 strips, dilute 25 µl of the stock solution in 5 ml of buffer. Partially used stock reagents may be kept in their original vials and stored at -20°C for use in a future assay.

NOTE 7: This kit contains animal components from non-infectious animals and should be treated as potential biohazards in use and for disposal.

NOTE 8: Depending on the isotypes, subtypes, and targeting epitopes of antibodies, the binding affinity of individual antibodies varies significantly. For example, the total CPE IgG antibody concentration calculated as the sum of individual anti-CPE IgG subtypes might not perfectly match the total IgG concentration as determined by the Mouse Anti-Ara h2 IgG Antibody ELISA Kit (Cat # 3077).

ASSAY PROCEDURE

- Add Blocking Buffer:** Add 100 μl of the Blocking Buffer (Solution B) to each well and incubate at room temperature for 1 hour.
- Prepare Standard Dilutions:** Please see the figure below for each assay's recommended standard range. Dissolve one vial of Standard in 1 ml of Sample/Standard/Secondary Antibody Dilution Buffer (Solution C) and keep it as a standard stock. Dilute the standard stock accordingly to get the first stock solution. Then serially dilute it with Solution C. For example, mix 250 μl of the first stock solution with an equal volume of Solution C to make the second stock solution, and then repeat it five more times. The original standard stock solution may be stored at -20°C for use in a future assay. Chondrex, Inc. recommends making fresh serial dilutions for each assay.



- Prepare Sample Dilutions:** An important point to note is that the composition of CPE mixtures can exhibit variations depending on the vendor, batch, and manufacturing process. These variations can result in differing levels of antigens in the final CPE product, which when used to immunize mice, can impact the serum antibody levels against those antigens. The dilution of serum from mice immunized with CPE antigens varies (1:100 or more) depending on the immunization schedule and timing of serum collection. In general, no antibodies against CPE are observed in normal serum at a 1:100 dilution. If serum samples require a lower dilution than 1:100, please contact support@chondrex.com.
- Wash:** Dilute 50 ml of 20X wash buffer in 950 ml of distilled water (1X wash buffer). Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blotting on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
- Add Standards and Samples:** Add 100 μl of standards, Solution C (blank), and samples to wells in duplicate. Incubate at room temperature for 2 hours.
- Wash:** Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blotting on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
- Add Secondary Antibody:** Dilute one vial of Secondary Antibody in 10 ml Sample/Standard/Secondary Antibody Dilution Buffer (Solution C). Add 100 μl of secondary antibody solution to each well and incubate at room temperature for 1 hour.

Strip #	2 nd Antibody (μl)	Solution C (ml)
2	8	1.7
4	17	3.3
6	25	5.0
8	33	6.6
10	42	8.2
12	50	10.0

8. **Wash:** Wash the plate with 1X wash buffer at least 3 times using a wash bottle with manifold or an automated plate washer. Empty the plate by inverting it and blotting on a paper towel to remove excess liquid. *Do not allow the plate to dry out.*
9. **Add TMB Solution:** Use new tubes when preparing TMB solution. Just prior to use, prepare TMB solution with Chromogen Dilution Buffer as shown in the following table. Add 100 μ l of TMB solution to each well immediately after washing the plate and incubate for 25 minutes at room temperature.


Strip #	TMB (μ l)	Chromogen Dilution Buffer (ml)
2	34	1.7
4	66	3.3
6	100	5.0
8	132	6.6
10	164	8.2
12	200	10.0


10. **Stop:** Stop the reaction with 50 μ l of 2N Sulfuric Acid (Stop Solution) to each well.
11. **Read Plate:** Read the OD values at 450 nm. If the OD values of samples are greater than the OD values of the highest standard, re-assay the samples at a higher dilution. A 630 nm filter can be used as a reference.

CALCULATING RESULTS

1. Average the duplicate OD values for the standards, blanks (B), and test samples.
2. Subtract the averaged blank OD values from the averaged OD values of the standards and test samples.
3. Plot the OD values of the standards against the ng/ml of antibody standard. Using a log/log plot will linearize the data. Figure 1 shows sample standard curves for anti-Ara h2 Ig antibodies.
4. The ng/ml of antibody in test samples can be calculated using regression analysis. Multiply it by the sample dilution factor to obtain the antibody concentration (ng/ml) in original test samples.

Figure 1 - Typical Standard Curves for the Anti-Ara h2 Antibody ELISA Kits





VALIDATION DATA

Table 1 - Reproducibility Data for the Mouse Anti-Ara h2 IgG Antibody ELISA Kit

Test	0.3 ng/ml	1.25 ng/ml	5 ng/ml
Intra-Assay CV (%)	6.4	6.1	8.8
Inter-Assay CV (%)	2.6	4.2	5.5
Spike Test* (%)	108%	99%	99%

Table 2 - Reproducibility Data for the Mouse Anti-Ara h2 IgG2b Antibody ELISA Kit

Test	0.3 ng/ml	1.25 ng/ml	5 ng/ml
Intra-Assay CV (%)	6.1	3.8	3.6
Inter-Assay CV (%)	5.5	9.4	3.2
Spike Test* (%)	109%	106%	108%

Table 3 - Reproducibility Data for the Mouse Anti-Ara h2 IgA Antibody ELISA Kit

Test	6.3 ng/ml	25 ng/ml	100 ng/ml
Intra-Assay CV (%)	1.9	3.8	5.2
Inter-Assay CV (%)	2.2	6.9	1.5
Spike Test* (%)	103%	96%	94%

Table 4 - Reproducibility Data for the Mouse Anti-Ara h2 IgM Antibody ELISA Kit

Test	6.3 ng/ml	25 ng/ml	100 ng/ml
Intra-Assay CV (%)	5.2	5.6	3.0
Inter-Assay CV (%)	6.7	8.4	7.1
Spike Test* (%)	110%	100%	110%

*Known amounts of anti-Ara h2 antibodies were added to samples and then diluted with Sample/Standard/Secondary Antibody Dilution Buffer to assay anti-Ara h2 antibodies by ELISA.

TROUBLESHOOTING

For frequently asked questions about assays and ELISAs, please see Chondrex, Inc.'s [ELISA FAQ](#) for more information.

REFERENCES

1. A. W. Burks, Peanut allergy. *Lancet*. **371**, 1538–1546 (2008).
2. Y. Zhuang, S. C. Dreskin, Redefining the major peanut allergens. *Immunol. Res.* **55**, 125–134 (2013).
3. J. J. Dolence, Induction of Peanut Allergy Through Inhalation of Peanut in Mice. *Methods Mol. Biol.* **2223**, 19–35 (2021).
4. X. M. Li, D. Serebrisky, S. Y. Lee, C. K. Huang, L. Bardina, B. H. Schofield, J. S. Stanley, A. W. Burks, G. A. Bannon, H. A. Sampson, A murine model of peanut anaphylaxis: T- and B-cell responses to a major peanut allergen mimic human responses. *J. Allergy Clin. Immunol.* **106**, 150–158 (2000).
5. M.-J. Bae, H. S. Shin, E.-K. Kim, J. Kim, D.-H. Shon, Oral administration of chitin and chitosan prevents peanut-induced anaphylaxis in a murine food allergy model. *Int. J. Biol. Macromol.* **61**, 164–168 (2013).
6. M. Kulis, X. Chen, J. Lew, Q. Wang, O. P. Patel, Y. Zhuang, K. S. Murray, M. W. Duncan, H. S. Porterfield, A. W. Burks, S. C. Dreskin, The 2S albumin allergens of *Arachis hypogaea*, Ara h 2 and Ara h 6, are the major elicitors of anaphylaxis and can effectively desensitize peanut-allergic mice. *Clin. Exp. Allergy*. **42**, 326–336 (2012).
7. L. M. Chang, Y. Song, X.-M. Li, H. A. Sampson, M. Masilamani, Dietary Elimination of Soybean Components Enhances Allergic Immune Response to Peanuts in BALB/c Mice. *Int. Arch. Allergy Immunol.* **166**, 304–310 (2015).
8. C. Zhou, T. Ludmila, N. Sun, C. Wang, Q. Pu, K. Huang, H. Che, BALB/c mice can be used to evaluate allergenicity of different food protein extracts. *Food Agric. Immunol.* **27**, 589–603 (2016).
9. C. C. Bowman, M. K. Selgrade, Differences in allergenic potential of food extracts following oral exposure in mice reflect differences in digestibility: potential approaches to safety assessment. *Toxicol. Sci.* **102**, 100–109 (2008).